

PHILIP MORRIS U.S.A.
INTEROFFICE CORRESPONDENCE
Richmond, Virginia

To: Distribution **Date:** October 3, 1997
From: J. M. Garman
Subject: Massachusetts Regulations: Cigarettes / Smoking Machine Port Study

The puff parameters documented within the Massachusetts Regulations produce a significant increase in the delivery of Total Particulate Matter within the smoking analysis. Bleeding of TPM material through the Cambridge filter pad occurs when applying the FTC practice of smoking 5 cigarettes per smoking machine port. The FTC procedure does not specify a maximum limit of TPM on a 44 mm Cambridge filter pad. ISO 4387, "Cigarettes - Determination of Total and Nicotine Free Dry Particulate Matter using a Routine Analytical Smoking Machine," 1991, states that the amount of TPM on a 44 mm Cambridge filter pad shall not exceed 150 milligrams. Utilizing the Massachusetts Regulation, TPM's of greater than 300 milligrams per filter pad are possible when smoking 5 high-delivery cigarettes through one Cambridge filter pad.

In this study, the Massachusetts and normal FTC smoking procedures were applied to a non-filtered brand, (Best Buy, Brand Code BAZ1), where 16 ports of data was obtained via three and via five cigarettes per port. Data generated as part of a taping study, J. M. Garman memo, "Massachusetts Regulations: Cigarette Taping Study," October 2, 1997, was also used in this study.

All the resultant Cambridge filter pads did show bleed-through of TPM material. The attached results show that, at the TPM levels measured, this bleed-through has no effect on the resultant nicotine and tar results. The results for five cigarettes per port were consistent to those obtained utilizing three cigarettes per port.

Distribution

L. Chambers
C. Connell
R. Cox
C. Hendrick
K. Hughes
K. Podraza
B. Strang
J. Whidby

**MASSACHUSETTS REGULATION:
NUMBER OF CIGARETTES PER PORT STUDY**

Sample	Puff Count	TPM mg/cigt	Nicotine mg/cigt	Water mg/cigt	Tar mg/cigt	CO mg/cigt
2R1F (5 cigt/port) *	13.8	56.5	2.59	18.52	35.4	32.7
Std. Dev.	0.35	1.81	0.115	1.20	0.98	1.21
N	16	16	16	16	16	16
2R1F (3 cigt/port) **	14.2	57.4	2.54	20.06	34.8	31.7
Std. Dev.	0.45	3.99	0.139	2.27	1.79	1.63
N	15	15	15	15	15	13
1R3F (5 cigt/port) *	12.1	45.9	2.25	14.25	29.4	26.1
Std. Dev.	0.34	1.62	0.099	0.97	0.94	0.93
N	16	16	16	16	16	16
1R3F (3 cigt/port) **	12.6	50.5	2.25	17.96	30.3	27.5
Std. Dev.	0.75	1.65	0.118	1.30	1.11	0.91
N	15	15	15	15	15	12
IM#16 (5 cigt/port) *	10.6	45.2	1.86	14.27	29.1	26.3
Std. Dev.	0.29	1.74	0.052	0.90	1.06	0.93
N	16	16	16	16	16	16
IM#16 (3 cigt/port) *	10.7	45.5	1.83	14.67	28.9	26.5
Std. Dev.	0.42	2.19	0.155	1.05	1.18	0.87
N	16	15	15	15	15	14
BAZ1 (5 cigt/port) ***	13.0	63.0	2.96	12.14	47.9	26.2
Std. Dev.	0.38	1.27	0.057	0.71	0.65	0.72
N	16	16	16	16	16	16
BAZ1 (3 cigt/port) ***	13.2	63.1	3.00	12.40	47.7	26.5
Std. Dev.	0.45	2.47	0.165	0.87	1.64	0.98
N	16	16	16	16	16	16

* Data obtained from study documented in memo by J. M. Garman, "Massachusetts Regulations: Cigarette Taping Study," October 2, 1997.

** Data obtained as part of above study. Cigarette tipping paper completely taped.

*** Best Buy Non-Filter. Cigarettes unequilibrated.